

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045010.D
 Acq On : 29 Sep 2021 21:57
 Operator : SY/MD
 Sample : M3926-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EP-1-MW-3R-0921

Quant Time: Sep 30 04:34:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	213242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	383208	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	356219	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	171336	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	167277	47.12	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.24%		
35) Dibromofluoromethane	5.295	113	117714	48.69	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.38%		
50) Toluene-d8	7.902	98	482605	50.53	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.06%		
62) 4-Bromofluorobenzene	10.635	95	180626	45.80	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.60%		
Target Compounds						Qvalue
39) Methylcyclohexane	6.767	83	14500	2.95	ug/l	99
67) Ethyl Benzene	9.574	91	509361	37.06	ug/l	99
68) m/p-Xylenes	9.697	106	431003	80.72	ug/l	99
69) o-Xylene	10.102	106	74170	14.00	ug/l	100
73) Isopropylbenzene	10.488	105	964169	75.57	ug/l	100
78) n-propylbenzene	10.909	91	1745810	112.28	ug/l	100
80) 1,3,5-Trimethylbenzene	11.092	105	348431	31.17	ug/l	100
83) tert-Butylbenzene	11.423	119	104640	9.62	ug/l	89
84) 1,2,4-Trimethylbenzene	11.475	105	7920814	716.28	ug/l	100
85) sec-Butylbenzene	11.648	105	531081	37.93	ug/l	100
89) n-Butylbenzene	12.211	91	287364	27.70	ug/l #	79
95) Naphthalene	14.089	128	81096	6.41	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045010.D
Acq On : 29 Sep 2021 21:57
Operator : SY/MD
Sample : M3926-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EP-1-MW-3R-0921

Quant Time: Sep 30 04:34:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

