

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045014.D
 Acq On : 30 Sep 2021 00:19
 Operator : SY/MD
 Sample : VU0929WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0929WBL02

Quant Time: Sep 30 00:37:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	210739	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	372734	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	351005	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	154271	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	164801	46.98	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.96%
35) Dibromofluoromethane	5.292	113	114242	48.59	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.18%
50) Toluene-d8	7.899	98	465687	50.13	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.26%
62) 4-Bromofluorobenzene	10.636	95	171140	44.62	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.24%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045014.D
Acq On : 30 Sep 2021 00:19
Operator : SY/MD
Sample : VU0929WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0929WBL02

Quant Time: Sep 30 00:37:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

