

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045020.D
 Acq On : 30 Sep 2021 02:40
 Operator : SY/MD
 Sample : PB139450ZHE#03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB139450ZHE#03

Quant Time: Sep 30 05:29:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	196841	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	357466	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	359308	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	165494	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	154805	47.24	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.48%
35) Dibromofluoromethane	5.292	113	107515	47.68	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.36%
50) Toluene-d8	7.903	98	456395	51.22	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.44%
62) 4-Bromofluorobenzene	10.636	95	183367	49.85	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.70%
Target Compounds						
						Qvalue
16) Acetone	2.629	43	47946	24.65	ug/l	100
18) Methyl Acetate	2.957	43	7951	1.40	ug/l	95
20) Methylene Chloride	3.047	84	12485	4.26	ug/l	96
43) Isopropyl Acetate	5.915	43	26204	3.52	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045020.D
Acq On : 30 Sep 2021 02:40
Operator : SY/MD
Sample : PB139450ZHE#03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB139450ZHE#03

Quant Time: Sep 30 05:29:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

