

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045022.D
 Acq On : 30 Sep 2021 03:27
 Operator : SY/MD
 Sample : M3952-05DL 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-3DL

Quant Time: Sep 30 05:29:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	206585	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	371616	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	351970	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	168979	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	163613	47.58	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.16%		
35) Dibromofluoromethane	5.292	113	115473	49.26	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.52%		
50) Toluene-d8	7.902	98	476126	51.40	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	102.80%		
62) 4-Bromofluorobenzene	10.635	95	180097	47.10	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	94.20%		
Target Compounds						Qvalue
16) Acetone	2.629	43	5211	2.55	ug/l	99
39) Methylcyclohexane	6.764	83	140761	29.50	ug/l	98
40) Benzene	5.777	78	1907622	164.81	ug/l	100
52) Toluene	7.973	92	43787	6.04	ug/l	99
67) Ethyl Benzene	9.574	91	93273	6.87	ug/l	99
68) m/p-Xylenes	9.697	106	162122	30.73	ug/l	98
69) o-Xylene	10.102	106	4562	0.87	ug/l	96
73) Isopropylbenzene	10.488	105	23348	1.86	ug/l	100
78) n-propylbenzene	10.909	91	34151	2.23	ug/l	98
80) 1,3,5-Trimethylbenzene	11.092	105	53215	4.83	ug/l	100
83) tert-Butylbenzene	11.423	119	3231	0.30	ug/l	86
84) 1,2,4-Trimethylbenzene	11.471	105	158004	14.49	ug/l	100
85) sec-Butylbenzene	11.645	105	9670	0.70	ug/l	95
86) p-Isopropyltoluene	11.796	119	8756	0.79	ug/l	90
89) n-Butylbenzene	12.208	91	11784	1.15	ug/l #	41
95) Naphthalene	14.089	128	38245	3.07	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045022.D
Acq On : 30 Sep 2021 03:27
Operator : SY/MD
Sample : M3952-05DL 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-3DL

Quant Time: Sep 30 05:29:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

