

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045024.D
 Acq On : 30 Sep 2021 04:14
 Operator : SY/MD
 Sample : M3952-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-1

Quant Time: Oct 01 11:50:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	206133	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	370253	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	344275	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	153144	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	163931	47.77	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.54%
35) Dibromofluoromethane	5.295	113	115029	49.25	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.50%
50) Toluene-d8	7.902	98	462752	50.14	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.28%
62) 4-Bromofluorobenzene	10.635	95	163871	43.01	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.02%
Target Compounds						
						Qvalue
18) Methyl Acetate	2.957	43	3196	0.54	ug/l	96
27) cis-1,2-Dichloroethene	4.671	96	5104	1.62	ug/l	86
44) Trichloroethene	6.549	130	2660	0.98	ug/l	93
64) Tetrachloroethene	8.555	164	276839	123.20	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045024.D
Acq On : 30 Sep 2021 04:14
Operator : SY/MD
Sample : M3952-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-1

Quant Time: Oct 01 11:50:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

