

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045026.D
 Acq On : 30 Sep 2021 05:01
 Operator : SY/MD
 Sample : M3952-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-2

Quant Time: Sep 30 05:30:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	206412	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	368918	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	346047	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	152002	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	167437	48.73	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.46%
35) Dibromofluoromethane	5.295	113	116459	50.04	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.08%
50) Toluene-d8	7.903	98	466413	50.72	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.44%
62) 4-Bromofluorobenzene	10.636	95	165821	43.68	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.36%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045026.D
Acq On : 30 Sep 2021 05:01
Operator : SY/MD
Sample : M3952-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-2

Quant Time: Sep 30 05:30:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

