

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045027.D
 Acq On : 30 Sep 2021 05:25
 Operator : SY/MD
 Sample : M3952-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-4

Quant Time: Sep 30 05:54:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	204875	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	364688	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	340646	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	151275	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	164618	48.27	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.54%		
35) Dibromofluoromethane	5.295	113	115235	50.09	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.18%		
50) Toluene-d8	7.902	98	462069	50.83	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.66%		
62) 4-Bromofluorobenzene	10.635	95	163767	43.64	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	87.28%		
Target Compounds						
27) cis-1,2-Dichloroethene	4.674	96	2471	0.79	ug/l	82
64) Tetrachloroethene	8.555	164	6882	3.10	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045027.D
Acq On : 30 Sep 2021 05:25
Operator : SY/MD
Sample : M3952-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-4

Quant Time: Sep 30 05:54:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

