

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045029.D
 Acq On : 30 Sep 2021 06:12
 Operator : SY/MD
 Sample : M3971-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RO-COMP-1

Quant Time: Sep 30 06:44:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	203221	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	361461	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	350260	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	157242	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	164958	48.76	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.52%
35) Dibromofluoromethane	5.295	113	115568	50.68	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.36%
50) Toluene-d8	7.903	98	471042	52.28	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.56%
62) 4-Bromofluorobenzene	10.636	95	168541	45.31	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.62%
Target Compounds						
16) Acetone	2.630	43	32948	16.41	ug/l	98
20) Methylene Chloride	3.051	84	9570	3.16	ug/l	94
43) Isopropyl Acetate	5.916	43	23680	3.15	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045029.D
Acq On : 30 Sep 2021 06:12
Operator : SY/MD
Sample : M3971-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RO-COMP-1

Quant Time: Sep 30 06:44:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

