

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045031.D
 Acq On : 30 Sep 2021 06:59
 Operator : SY/MD
 Sample : M3971-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RP-COMP-3

Quant Time: Sep 30 08:29:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	200445	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	360434	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	360236	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	163477	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	160249	48.03	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.06%
35) Dibromofluoromethane	5.292	113	110818	48.74	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.48%
50) Toluene-d8	7.902	98	468575	52.16	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.32%
62) 4-Bromofluorobenzene	10.635	95	179861	48.49	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.98%
Target Compounds						
					Qvalue	
16) Acetone	2.629	43	37672	19.02	ug/l	100
18) Methyl Acetate	2.954	43	6680	1.15	ug/l	96
20) Methylene Chloride	3.047	84	10414	3.49	ug/l	92
43) Isopropyl Acetate	5.915	43	22334	2.98	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045031.D
Acq On : 30 Sep 2021 06:59
Operator : SY/MD
Sample : M3971-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RP-COMP-3

Quant Time: Sep 30 08:29:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

