


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045012.D
Acq On : 29 Sep 2021 23:08
Operator : SY/MD
Sample : BFB
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Title : SW846 8260
Last Update : Tue Sep 28 12:32:13 2021



AutoFind: Scans 2890, 2891, 2892; Background Corrected with Scan 2878

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	8933	PASS
75	95	30	60	54.1	22821	PASS
95	95	100	100	100.0	42176	PASS
96	95	5	9	6.0	2518	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	70.5	29741	PASS
175	174	5	9	7.3	2176	PASS
176	174	95	101	96.7	28752	PASS
177	176	5	9	6.0	1739	PASS