

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU093020\
 Data File : VU040391.D
 Acq On : 30 Sep 2020 12:10
 Operator : SY/MD
 Sample : L4186-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG0E2

Quant Time: Oct 01 03:16:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 01 03:05:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	125830	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	128583	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	58649	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	28135	4.52	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.40%
7) Chloroethane-d5	1.92	69	29255	5.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.20%
11) 1,1-Dichloroethene-d2	2.58	63	46547	3.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.20%
20) 2-Butanone-d5	4.65	46	162551	52.53	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.06%
24) Chloroform-d	5.08	84	75226	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.40%
26) 1,2-Dichloroethane-d4	5.72	65	47301	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.80%
32) Benzene-d6	5.74	84	130290	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.80%
36) 1,2-Dichloropropane-d6	6.70	67	48209	5.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.00%
41) Toluene-d8	7.91	98	100516	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
43) trans-1,3-Dichloropropene-	8.19	79	18474	4.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.80%
46) 2-Hexanone-d5	8.64	63	114199	50.31	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.62%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	45647	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	46679	5.25	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.00%

Target Compounds

					Ovalue
5) Vinyl chloride	1.61	62	1558	0.128 ug/L #	52
13) Acetone	2.66	43	4572	1.773 ug/L	94
14) Carbon disulfide	2.81	76	1578	0.052 ug/L #	93
18) trans-1,2-Dichloroethene	3.37	96	1859	0.188 ug/L	92
22) cis-1,2-Dichloroethene	4.68	96	124042	11.316 ug/L	97
42) Toluene	7.97	91	1550	0.035 ug/L #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU093020\
Data File : VU040391.D
Acq On : 30 Sep 2020 12:10
Operator : SY/MD
Sample : L4186-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG0E2

Quant Time: Oct 01 03:16:06 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 01 03:05:40 2020
Response via : Initial Calibration

