

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU093020\
 Data File : VU040401.D
 Acq On : 30 Sep 2020 16:03
 Operator : SY/MD
 Sample : VIBLK65
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK65

Quant Time: Oct 01 03:57:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 01 03:05:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	132692	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	136414	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	60649	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	26098	3.97	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.40%
7) Chloroethane-d5	1.92	69	25074	4.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.20%
11) 1,1-Dichloroethene-d2	2.58	63	43218	2.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.80%#
20) 2-Butanone-d5	4.66	46	156240	47.88	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.76%
24) Chloroform-d	5.08	84	68194	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
26) 1,2-Dichloroethane-d4	5.72	65	43844	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
32) Benzene-d6	5.75	84	121496	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.60%
36) 1,2-Dichloropropane-d6	6.70	67	45074	4.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.80%
41) Toluene-d8	7.91	98	94182	3.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.00%
43) trans-1,3-Dichloropropene-	8.19	79	16589	4.05	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.00%
46) 2-Hexanone-d5	8.64	63	107894	44.80	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.60%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	43955	4.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	45296	4.93	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.60%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	3215	0.253 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU093020\
Data File : VU040401.D
Acq On : 30 Sep 2020 16:03
Operator : SY/MD
Sample : VIBLK65
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK65

Quant Time: Oct 01 03:57:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 01 03:05:40 2020
Response via : Initial Calibration

