

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU093021\
 Data File : VU045036.D
 Acq On : 30 Sep 2021 10:53
 Operator : SY/MD
 Sample : VU0930MBL01
 Misc : 5.0g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0930MBL01

Quant Time: Sep 30 16:06:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	198604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	360579	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	330196	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	141651	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	161711	48.914	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.820%
35) Dibromofluoromethane	5.295	113	113516	49.905	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.800%
50) Toluene-d8	7.902	98	449143	49.974	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.940%
62) 4-Bromofluorobenzene	10.635	95	156281	42.118	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU093021\
Data File : VU045036.D
Acq On : 30 Sep 2021 10:53
Operator : SY/MD
Sample : VU0930MBL01
Misc : 5.0g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0930MBL01

Quant Time: Sep 30 16:06:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

