

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU093021\
 Data File : VU045046.D
 Acq On : 30 Sep 2021 14:53
 Operator : SY/MD
 Sample : M3966-06
 Misc : 10.75g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S-833-K2-SO-12-14-092721

Quant Time: Oct 01 05:39:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	211773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	389030	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.427	117	364717	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	163692	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	174550	49.515	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.020%
35) Dibromofluoromethane	5.292	113	117740	47.976	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.960%
50) Toluene-d8	7.902	98	490482	50.582	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.160%
62) 4-Bromofluorobenzene	10.642	95	182558	45.602	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU093021\
Data File : VU045046.D
Acq On : 30 Sep 2021 14:53
Operator : SY/MD
Sample : M3966-06
Misc : 10.75g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S-833-K2-SO-12-14-092721

Quant Time: Oct 01 05:39:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

