

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU093021\
 Data File : VU045047.D
 Acq On : 30 Sep 2021 15:17
 Operator : SY/MD
 Sample : M3937-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Oct 01 05:39:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	195691	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	356882	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	342313	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	153446	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	159861	49.074	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.140%
35) Dibromofluoromethane	5.292	113	109341	48.567	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.140%
50) Toluene-d8	7.899	98	452256	50.842	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.680%
62) 4-Bromofluorobenzene	10.635	95	172184	46.885	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU093021\
Data File : VU045047.D
Acq On : 30 Sep 2021 15:17
Operator : SY/MD
Sample : M3937-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Oct 01 05:39:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

