

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU093021\
 Data File : VU045052.D
 Acq On : 30 Sep 2021 17:14
 Operator : SY/MD
 Sample : M3937-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 01 05:40:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	193474	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	351407	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	320967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	145792	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	161500	50.146	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.300%
35) Dibromofluoromethane	5.295	113	111613	50.349	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.700%
50) Toluene-d8	7.902	98	441218	50.374	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.740%
62) 4-Bromofluorobenzene	10.635	95	154021	42.593	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU093021\
Data File : VU045052.D
Acq On : 30 Sep 2021 17:14
Operator : SY/MD
Sample : M3937-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 01 05:40:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

