

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100118\
 Data File : VU027320.D
 Acq On : 01 Oct 2018 14:58
 Operator : MD/SY
 Sample : VU1001WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1001WBL01

Quant Time: Oct 02 07:42:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	103710	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	174788	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	164306	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	77336	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	100313	57.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.84%	
35) Dibromofluoromethane	4.88	113	71543	59.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	119.44%	
50) Toluene-d8	7.57	98	274808	61.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	122.60%	
62) 4-Bromofluorobenzene	10.31	95	89092	53.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.94%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100118\
Data File : VU027320.D
Acq On : 01 Oct 2018 14:58
Operator : MD/SY
Sample : VU1001WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1001WBL01

Quant Time: Oct 02 07:42:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

