

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100118\
 Data File : VU027330.D
 Acq On : 01 Oct 2018 19:56
 Operator : MD/SY
 Sample : J5149-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WS-3

Quant Time: Oct 02 08:07:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	102933	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	175150	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	167532	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	80265	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	103089	59.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.96%	
35) Dibromofluoromethane	4.89	113	73220	61.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.00%	
50) Toluene-d8	7.57	98	277665	61.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.62%	
62) 4-Bromofluorobenzene	10.31	95	92571	55.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.88%	
Target Compounds						
16) Acetone	2.33	43	6827	6.723	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100118\
Data File : VU027330.D
Acq On : 01 Oct 2018 19:56
Operator : MD/SY
Sample : J5149-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WS-3

Quant Time: Oct 02 08:07:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

