

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100118\
 Data File : VU027333.D
 Acq On : 01 Oct 2018 21:06
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 02 08:13:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.721	0.763	-5.8	89	0.00
3 P	Chloromethane	0.976	0.916	6.1	84	0.00
4 C	Vinyl Chloride	0.917	0.924	-0.8#	90	0.00
5 T	Bromomethane	0.397	0.366	7.8	89	0.00
6 T	Chloroethane	0.551	0.514	6.7	89	0.00
7 T	Trichlorofluoromethane	0.931	0.961	-3.2	93	0.00
8 T	Diethyl Ether	0.429	0.430	-0.2	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.541	0.554	-2.4	92	0.00
10 T	Methyl Iodide	0.394	0.318	19.3	70	0.00
11 T	Tert butyl alcohol	0.176	0.177	-0.6	96	0.00
12 CM	1,1-Dichloroethene	0.528	0.520	1.5#	89	0.00
13 T	Acrolein	0.148	0.139	6.1	91	0.00
14 T	Allyl chloride	1.254	1.239	1.2	89	0.00
15 T	Acrylonitrile	0.458	0.488	-6.6	96	0.00
16 T	Acetone	0.493	0.431	12.6	81	0.00
17 T	Carbon Disulfide	1.905	1.792	5.9	87	0.00
18 T	Methyl Acetate	1.125	1.176	-4.5	98	0.00
19 T	Methyl tert-butyl Ether	2.116	2.154	-1.8	92	0.00
20 T	Methylene Chloride	0.687	0.669	2.6	91	0.00
21 T	trans-1,2-Dichloroethene	0.593	0.584	1.5	90	0.00
22 T	Diisopropyl ether	2.549	2.675	-4.9	94	0.00
23 T	Vinyl Acetate	2.226	2.369	-6.4	93	0.00
24 P	1,1-Dichloroethane	1.310	1.361	-3.9	93	0.00
25 T	2-Butanone	0.673	0.701	-4.2	92	0.00
26 T	2,2-Dichloropropane	1.077	0.988	8.3	85	0.00
27 T	cis-1,2-Dichloroethene	0.678	0.679	-0.1	91	0.00
28 T	Bromochloromethane	0.632	0.675	-6.8	94	0.00
29 T	Tetrahydrofuran	0.424	0.460	-8.5	96	0.00
30 C	Chloroform	1.164	1.209	-3.9#	93	0.00
31 T	Cyclohexane	1.532	1.271	17.0	89	0.00
32 T	1,1,1-Trichloroethane	0.981	1.001	-2.0	91	0.00
33 S	1,2-Dichloroethane-d4	0.835	0.816	2.3	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.343	0.337	1.7	98	0.00
36 T	1,1-Dichloropropene	0.543	0.536	1.3	90	0.00
37 T	Ethyl Acetate	0.702	0.748	-6.6	98	0.00
38 T	Carbon Tetrachloride	0.490	0.490	0.0	91	0.00
39 T	Methylcyclohexane	0.635	0.628	1.1	88	0.00
40 TM	Benzene	1.581	1.593	-0.8	92	0.00
41 T	Methacrylonitrile	0.397	0.422	-6.3	95	0.00
42 TM	1,2-Dichloroethane	0.587	0.609	-3.7	93	0.00
43 T	Isopropyl Acetate	1.090	1.171	-7.4	94	0.00
44 TM	Trichloroethene	0.353	0.350	0.8	90	0.00
45 C	1,2-Dichloropropane	0.456	0.475	-4.2#	93	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100118\
 Data File : VU027333.D
 Acq On : 01 Oct 2018 21:06
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 02 08:13:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.258	0.271	-5.0	93	0.00
47 T	Bromodichloromethane	0.535	0.550	-2.8	91	0.00
48 T	Methyl methacrylate	0.542	0.590	-8.9	94	0.00
49 T	1,4-Dioxane	0.009	0.010	-11.1	90	0.00
50 S	Toluene-d8	1.282	1.307	-2.0	96	0.00
51 T	4-Methyl-2-Pentanone	0.708	0.785	-10.9	97	0.00
52 CM	Toluene	0.936	0.962	-2.8#	92	0.00
53 T	t-1,3-Dichloropropene	0.627	0.643	-2.6	91	0.00
54 T	cis-1,3-Dichloropropene	0.674	0.692	-2.7	90	0.00
55 T	1,1,2-Trichloroethane	0.371	0.384	-3.5	92	0.00
56 T	Ethyl methacrylate	0.616	0.676	-9.7	92	0.00
57 T	1,3-Dichloropropane	0.720	0.738	-2.5	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.344	0.379	-10.2	94	0.00
59 T	2-Hexanone	0.549	0.599	-9.1	95	0.00
60 T	Dibromochloromethane	0.372	0.388	-4.3	93	0.00
61 T	1,2-Dibromoethane	0.395	0.405	-2.5	92	0.00
62 S	4-Bromofluorobenzene	0.477	0.483	-1.3	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.311	0.305	1.9	89	0.00
65 PM	Chlorobenzene	1.052	1.021	2.9	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.371	-2.2	93	0.00
67 C	Ethyl Benzene	1.865	1.886	-1.1#	90	0.00
68 T	m/p-Xylenes	0.682	0.698	-2.3	90	0.00
69 T	o-Xylene	0.648	0.683	-5.4	91	0.00
70 T	Styrene	1.064	1.120	-5.3	90	0.00
71 P	Bromoform	0.292	0.297	-1.7	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.631	3.682	-1.4	90	0.00
74 T	N-amyl acetate	2.079	2.211	-6.3	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.426	1.439	-0.9	95	0.00
76 T	1,2,3-Trichloropropane	1.296	1.298	-0.2	92	0.00
77 T	Bromobenzene	0.844	0.839	0.6	92	0.00
78 T	n-propylbenzene	4.364	4.455	-2.1	90	0.00
79 T	2-Chlorotoluene	2.635	2.613	0.8	91	0.00
80 T	1,3,5-Trimethylbenzene	2.957	3.058	-3.4	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.495	0.382	22.8#	76	0.00
82 T	4-Chlorotoluene	2.990	3.009	-0.6	90	0.00
83 T	tert-Butylbenzene	2.858	2.904	-1.6	89	0.00
84 T	1,2,4-Trimethylbenzene	3.020	3.149	-4.3	91	0.00
85 T	sec-Butylbenzene	3.601	3.713	-3.1	90	0.00
86 T	p-Isopropyltoluene	3.083	3.184	-3.3	90	0.00
87 T	1,3-Dichlorobenzene	1.598	1.532	4.1	89	0.00
88 T	1,4-Dichlorobenzene	1.633	1.556	4.7	90	0.00
89 T	n-Butylbenzene	2.876	2.966	-3.1	88	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100118\
Data File : VU027333.D
Acq On : 01 Oct 2018 21:06
Operator : MD/SY
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Oct 02 08:13:50 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.581	0.579	0.3	91	0.00
91 T	1,2-Dichlorobenzene	1.556	1.536	1.3	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.316	0.327	-3.5	93	0.00
93 T	1,2,4-Trichlorobenzene	1.007	1.018	-1.1	89	0.00
94 T	Hexachlorobutadiene	0.509	0.487	4.3	88	0.00
95 T	Naphthalene	3.375	3.573	-5.9	91	0.00
96 T	1,2,3-Trichlorobenzene	1.049	1.043	0.6	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6