

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100120\
 Data File : VU040430.D
 Acq On : 01 Oct 2020 19:24
 Operator : SY/MD
 Sample : VU1001WBL02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK20

Quant Time: Oct 02 05:34:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 02 05:20:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	121252	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	124018	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	53999	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	33024	5.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.00%
7) Chloroethane-d5	1.92	69	26290	5.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.00%
11) 1,1-Dichloroethene-d2	2.58	63	49314	3.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.40%
20) 2-Butanone-d5	4.66	46	122239	40.99	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.98%
24) Chloroform-d	5.08	84	63805	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.40%
26) 1,2-Dichloroethane-d4	5.72	65	40563	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
32) Benzene-d6	5.75	84	125294	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
36) 1,2-Dichloropropane-d6	6.70	67	41689	4.57	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.40%
41) Toluene-d8	7.91	98	104328	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.40%
43) trans-1,3-Dichloropropene-	8.19	79	16870	4.53	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.60%
46) 2-Hexanone-d5	8.64	63	81099	37.04	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.08%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	35420	3.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	40524	4.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.00%

Target Compounds					Ovalue
16) Methylene chloride	3.07	84	2535	0.219 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100120\
Data File : VU040430.D
Acq On : 01 Oct 2020 19:24
Operator : SY/MD
Sample : VU1001WBL02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK20

Quant Time: Oct 02 05:34:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 02 05:20:15 2020
Response via : Initial Calibration

