

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100120\  
 Data File : VU040435.D  
 Acq On : 01 Oct 2020 21:28  
 Operator : SY/MD  
 Sample : L4188-06  
 Misc : 25.0mL/MSVOA U/WATER  
 ALS Vial : 27 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VHBLK01

Quant Time: Oct 02 05:53:28 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR092920WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Fri Oct 02 05:20:15 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 6.26  | 114  | 118777   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 9.42  | 117  | 119012   | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.81 | 152  | 51146    | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.60   | 65    | 35285    | 6.00     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 120.00% |
| 7) Chloroethane-d5             | 1.92   | 69    | 28450    | 5.53     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 110.60% |
| 11) 1,1-Dichloroethene-d2      | 2.58   | 63    | 54108    | 3.83     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 76.60%  |
| 20) 2-Butanone-d5              | 4.66   | 46    | 135643   | 46.44    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 92.88%  |
| 24) Chloroform-d               | 5.08   | 84    | 68004    | 4.86     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 97.20%  |
| 26) 1,2-Dichloroethane-d4      | 5.72   | 65    | 44699    | 5.24     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 104.80% |
| 32) Benzene-d6                 | 5.75   | 84    | 136306   | 5.70     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 114.00% |
| 36) 1,2-Dichloropropane-d6     | 6.70   | 67    | 46750    | 5.34     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 106.80% |
| 41) Toluene-d8                 | 7.91   | 98    | 111488   | 5.36     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 107.20% |
| 43) trans-1,3-Dichloropropene- | 8.19   | 79    | 17887    | 5.01     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 100.20% |
| 46) 2-Hexanone-d5              | 8.64   | 63    | 92372    | 43.97    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 87.94%  |
| 57) 1,1,2,2-Tetrachloroethane- | 10.75  | 84    | 40035    | 4.70     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 94.00%  |
| 64) 1,2-Dichlorobenzene-d4     | 12.19  | 152   | 43607    | 5.62     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 112.40% |

| Target Compounds       |      |    |      |            | Ovalue |
|------------------------|------|----|------|------------|--------|
| 16) Methylene chloride | 3.06 | 84 | 3433 | 0.302 ug/L | 97     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100120\  
Data File : VU040435.D  
Acq On : 01 Oct 2020 21:28  
Operator : SY/MD  
Sample : L4188-06  
Misc : 25.0mL/MSVOA U/WATER  
ALS Vial : 27 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VHBLK01

Quant Time: Oct 02 05:53:28 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR092920WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Fri Oct 02 05:20:15 2020  
Response via : Initial Calibration

