


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU100120\
Data File : VU040411.D
Acq On : 01 Oct 2020 09:06
Operator : SY/MD
Sample : BFB11
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB11

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
Title : TRACE VOA SOM01.0
Last Update : Fri Oct 02 05:20:15 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.7	51312	PASS
75	95	30	80	55.6	131533	PASS
95	95	100	100	100.0	236736	PASS
96	95	5	9	6.7	15784	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	73.7	174592	PASS
175	174	5	9	7.7	13429	PASS
176	174	95	101	96.1	167765	PASS
177	176	5	9	6.4	10755	PASS

m/z	Abundance
36.00	2820.0
37.10	13923.0
38.00	12968.0
39.10	5259.0
40.00	386.0
43.00	145.0
44.10	2268.0
45.00	2393.0
46.10	382.0
47.00	3350.0
48.00	1561.0
49.10	12144.0
50.10	52040.0
51.10	16832.0
52.10	648.0
53.00	122.0
55.00	615.0
56.10	4222.0
57.10	7752.0
58.00	253.0
60.00	2732.0
61.10	13435.0
62.10	13881.0
63.10	10546.0
64.00	1340.0
65.10	1323.0
66.10	208.0
67.10	708.0
68.10	26240.0
69.10	26888.0
70.00	2243.0
70.90	136.0
72.10	1717.0
73.10	11967.0
74.10	46192.0
75.10	134464.0
76.10	11592.0
77.10	1454.0
78.00	1065.0
79.00	10638.0
80.00	3076.0
81.00	10890.0
82.00	2486.0
82.90	324.0
86.00	413.0
87.00	9816.0
88.00	9651.0
90.30	133.0
91.00	1047.0
92.10	7465.0
93.10	11427.0
94.10	31880.0
95.10	236992.0
96.10	15973.0
97.10	519.0
102.80	183.0
104.00	1323.0
105.00	287.0
106.00	1244.0
106.90	228.0
109.90	217.0
110.90	284.0
112.00	164.0
112.90	272.0
114.90	333.0
115.90	1177.0
117.00	1878.0
118.00	1104.0
118.90	1706.0
122.90	103.0
124.00	106.0
125.00	135.0
125.90	155.0
128.00	898.0
129.00	421.0
130.00	1033.0
131.00	318.0
135.00	473.0
137.00	341.0
139.90	160.0
141.00	2705.0
142.00	481.0
143.00	3407.0
144.10	142.0
145.00	609.0
145.90	335.0
146.90	113.0
148.00	544.0
149.00	196.0
150.00	283.0
152.00	115.0
152.90	157.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11

154.00	208.0
155.00	558.0
157.10	333.0
158.90	263.0
161.00	242.0
170.20	104.0
170.90	214.0
172.10	1651.0
174.10	172224.0
175.10	13661.0
176.10	165248.0
177.10	10832.0
178.10	369.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11

m/z	Abundance
36.00	2849.0
37.10	13856.0
38.10	12556.0
39.10	5432.0
40.00	413.0
43.00	162.0
44.10	2106.0
45.00	2476.0
46.10	319.0
47.00	3247.0
48.00	1715.0
49.00	12221.0
50.10	53064.0
51.10	16928.0
52.10	642.0
53.00	102.0
55.00	654.0
56.00	4246.0
57.10	7951.0
58.00	267.0
60.00	2782.0
61.10	13339.0
62.10	14345.0
63.10	10584.0
64.00	1309.0
65.10	1222.0
66.00	179.0
67.10	745.0
68.10	26664.0
69.10	27656.0
70.10	2222.0
70.90	129.0
72.10	1719.0
73.10	12083.0
74.10	47608.0
75.10	136128.0
76.10	11591.0
77.10	1350.0
78.00	975.0
79.00	10850.0
80.00	2873.0
81.00	11364.0
81.90	2429.0
82.90	266.0
86.10	391.0
87.10	9914.0
88.00	9738.0
90.90	919.0
92.10	7623.0
93.10	11444.0
94.10	32160.0
95.10	244800.0
96.10	16408.0
97.10	485.0
102.80	191.0
104.00	1387.0
105.00	329.0
106.00	1287.0
106.90	200.0
109.90	238.0
110.90	273.0
112.00	218.0
112.90	278.0
115.00	313.0
115.90	1144.0
117.00	1855.0
118.00	1078.0
118.90	1560.0
123.90	108.0
125.00	110.0
125.90	137.0
128.00	899.0
129.00	383.0
130.00	1090.0
131.00	342.0
134.90	553.0
137.00	348.0
139.90	190.0
141.00	2568.0
142.00	404.0
143.00	3552.0
144.10	124.0
145.00	587.0
145.90	316.0
146.90	121.0
147.90	463.0
149.00	244.0
150.00	298.0
153.00	148.0
154.00	169.0
155.00	535.0
157.00	373.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11

158.90	311.0
161.00	266.0
170.90	192.0
172.10	1592.0
174.10	180928.0
175.10	13887.0
176.10	173632.0
177.10	11099.0
178.10	312.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11

m/z	Abundance
36.00	2545.0
37.10	12503.0
38.10	11500.0
39.10	4963.0
40.00	362.0
42.90	153.0
44.10	1786.0
45.00	2280.0
46.00	260.0
47.00	2769.0
48.00	1744.0
49.00	10989.0
50.10	48832.0
51.10	15244.0
52.10	559.0
55.00	635.0
56.00	4018.0
57.10	7302.0
58.00	283.0
60.00	2626.0
61.10	12271.0
62.10	12995.0
63.10	9624.0
64.00	1142.0
65.10	984.0
66.00	138.0
67.10	680.0
68.10	24272.0
69.10	25704.0
70.10	1967.0
72.10	1468.0
73.10	11062.0
74.10	44536.0
75.10	124008.0
76.10	10440.0
77.10	1198.0
78.00	799.0
79.00	10012.0
80.00	2487.0
81.00	10564.0
81.90	2248.0
83.00	208.0
86.00	303.0
87.00	9185.0
88.00	8895.0
91.00	840.0
92.10	7123.0
93.10	10324.0
94.10	29376.0
95.10	228416.0
96.10	14973.0
97.10	377.0
102.80	130.0
104.00	1274.0
105.00	324.0
106.00	1218.0
106.90	155.0
109.90	219.0
110.90	223.0
112.00	262.0
112.90	250.0
115.00	263.0
115.90	978.0
117.00	1682.0
117.90	1011.0
118.90	1321.0
128.00	846.0
129.00	318.0
130.00	928.0
131.00	332.0
134.90	555.0
137.00	289.0
139.80	184.0
141.00	2357.0
142.00	297.0
143.00	3277.0
144.10	107.0
145.10	473.0
145.90	302.0
146.90	108.0
147.90	440.0
149.00	248.0
150.00	269.0
153.00	147.0
153.90	137.0
155.00	488.0
157.00	422.0
159.00	324.0
161.00	251.0
171.00	155.0
172.10	1445.0
174.10	170624.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11

175.10	12741.0
176.10	164416.0
177.10	10334.0
178.10	232.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11

Scan 2875 (10.584 min): VU040411.D
BFB11

m/z	Abundance
39.80	128.0
43.90	402.0

Instrument :
MSVOA_U
ClientSampleId :
BFB11