

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100121\
 Data File : VU045069.D
 Acq On : 01 Oct 2021 13:42
 Operator : SY/MD
 Sample : PB139508TB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB139508TB

Quant Time: Oct 02 03:10:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	200656	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	366789	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	344161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	151674	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	166534	49.858	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.720%		
35) Dibromofluoromethane	5.295	113	115648	49.981	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.960%		
50) Toluene-d8	7.903	98	467929	51.183	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	102.360%		
62) 4-Bromofluorobenzene	10.636	95	168129	44.544	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.080%		
Target Compounds					Qvalue	
16) Acetone	2.639	43	32614	16.450	ug/l	98
20) Methylene Chloride	3.051	84	8301	2.780	ug/l	97
43) Isopropyl Acetate	5.919	43	25642	3.358	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100121\
Data File : VU045069.D
Acq On : 01 Oct 2021 13:42
Operator : SY/MD
Sample : PB139508TB
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB139508TB

Quant Time: Oct 02 03:10:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

