

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100121\
 Data File : VU045070.D
 Acq On : 01 Oct 2021 14:05
 Operator : SY/MD
 Sample : PB139508ZHE#01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB139508ZHE#01

Quant Time: Oct 02 03:11:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	201739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	376405	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	354630	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	151868	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	169639	50.515	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.020%
35) Dibromofluoromethane	5.292	113	115087	48.468	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.940%
50) Toluene-d8	7.899	98	471711	50.278	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.560%
62) 4-Bromofluorobenzene	10.639	95	174734	45.111	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.220%
Target Compounds						
					Qvalue	
16) Acetone	2.642	43	33297	16.704	ug/l	99
18) Methyl Acetate	2.961	43	6502	1.116	ug/l #	89
20) Methylene Chloride	3.047	84	8758	2.917	ug/l	91
43) Isopropyl Acetate	5.922	43	26575	3.392	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100121\
Data File : VU045070.D
Acq On : 01 Oct 2021 14:05
Operator : SY/MD
Sample : PB139508ZHE#01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB139508ZHE#01

Quant Time: Oct 02 03:11:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

