

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100121\
 Data File : VU045071.D
 Acq On : 01 Oct 2021 14:29
 Operator : SY/MD
 Sample : PB139508ZHE#02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB139508ZHE#02

Quant Time: Oct 02 03:11:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	199464	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	372141	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	354929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	154125	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	165180	49.748	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.500%		
35) Dibromofluoromethane	5.292	113	114104	48.605	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.200%		
50) Toluene-d8	7.902	98	470247	50.697	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.400%		
62) 4-Bromofluorobenzene	10.639	95	172913	45.153	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.300%		
Target Compounds					Qvalue	
16) Acetone	2.642	43	32158	16.317	ug/l	97
18) Methyl Acetate	2.957	43	6827	1.185	ug/l	100
20) Methylene Chloride	3.047	84	8565	2.885	ug/l	96
43) Isopropyl Acetate	5.919	43	26827	3.463	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100121\
Data File : VU045071.D
Acq On : 01 Oct 2021 14:29
Operator : SY/MD
Sample : PB139508ZHE#02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB139508ZHE#02

Quant Time: Oct 02 03:11:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

