

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100121\
 Data File : VU045074.D
 Acq On : 01 Oct 2021 15:38
 Operator : SY/MD
 Sample : M3999-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 26498

Quant Time: Oct 02 03:12:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	202724	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	378016	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	357123	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	152990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	166248	49.265	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.520%		
35) Dibromofluoromethane	5.292	113	115209	48.313	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.620%		
50) Toluene-d8	7.902	98	476337	50.555	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.120%		
62) 4-Bromofluorobenzene	10.639	95	177645	45.667	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.340%		
Target Compounds					Qvalue	
16) Acetone	2.642	43	17507	8.740	ug/l	98
18) Methyl Acetate	2.961	43	7282	1.244	ug/l	100
20) Methylene Chloride	3.047	84	13681	4.534	ug/l	93
25) 2-Butanone	4.722	43	20889	7.340	ug/l	94
43) Isopropyl Acetate	5.919	43	26880	3.416	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100121\
Data File : VU045074.D
Acq On : 01 Oct 2021 15:38
Operator : SY/MD
Sample : M3999-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
26498

Quant Time: Oct 02 03:12:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

