

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100121\
 Data File : VU045076.D
 Acq On : 01 Oct 2021 16:25
 Operator : SY/MD
 Sample : M3999-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 229-342

Quant Time: Oct 02 03:12:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	187321	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	353756	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	343265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	162433	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	158618	50.869	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.740%
35) Dibromofluoromethane	5.292	113	110700	49.605	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.220%
50) Toluene-d8	7.899	98	449112	50.934	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.860%
62) 4-Bromofluorobenzene	10.636	95	175605	48.239	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.480%
Target Compounds						
						Qvalue
16) Acetone	2.636	43	18229	9.849	ug/l	98
18) Methyl Acetate	2.957	43	7880	1.457	ug/l	95
20) Methylene Chloride	3.047	84	13693	4.912	ug/l	94
25) 2-Butanone	4.716	43	18161	6.906	ug/l	96
43) Isopropyl Acetate	5.919	43	22914	3.112	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100121\
Data File : VU045076.D
Acq On : 01 Oct 2021 16:25
Operator : SY/MD
Sample : M3999-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
229-342

Quant Time: Oct 02 03:12:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

