

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100121\
 Data File : VU045083.D
 Acq On : 01 Oct 2021 19:09
 Operator : SY/MD
 Sample : M4020-19
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB-01-093021

Quant Time: Oct 02 03:13:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.376	168	180551	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	335527	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	314751	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	146760	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.704	65	152864	50.861	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.720%	
35) Dibromofluoromethane	5.292	113	105542	49.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.720%	
50) Toluene-d8	7.900	98	426242	50.967	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.940%	
62) 4-Bromofluorobenzene	10.636	95	159904	46.312	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.620%	
Target Compounds						
						Qvalue
18) Methyl Acetate	2.958	43	3319	0.637	ug/l	# 68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100121\
Data File : VU045083.D
Acq On : 01 Oct 2021 19:09
Operator : SY/MD
Sample : M4020-19
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB-01-093021

Quant Time: Oct 02 03:13:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

