

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100124\
 Data File : VU061060.D
 Acq On : 01 Oct 2024 21:53
 Operator : MD/SY
 Sample : P4227-07 400X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Oct 02 02:38:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 02 02:09:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	175763	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	163594	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	76724	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	55304	4.401	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.000%
7) Chloroethane-d5	1.911	69	43323	4.606	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.200%
11) 1,1-Dichloroethene-d2	2.563	65	24197	4.365	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.400%
20) 2-Butanone-d5	4.631	46	148090	49.198	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.400%
24) Chloroform-d	5.055	84	100819	4.552	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
26) 1,2-Dichloroethane-d4	5.695	65	56406	4.805	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.200%
32) Benzene-d6	5.718	84	203932	4.728	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
36) 1,2-Dichloropropane-d6	6.682	67	69919	4.936	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.800%
41) Toluene-d8	7.891	98	182835	4.535	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.600%
43) trans-1,3-Dichloroprop...	8.177	79	27328	4.512	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.200%
46) 2-Hexanone-d5	8.627	63	121861	48.607	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.220%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	49810	4.795	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	59604	4.859	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.200%
Target Compounds						Qvalue
16) Methylene chloride	3.042	84	2081	0.179	ug/L	96
35) Methylcyclohexane	6.682	83	15381	0.729	ug/L #	19
37) 1,2-Dichloropropane	6.682	63	7513	0.591	ug/L #	93
51) Chlorobenzene	9.441	112	74050	2.366	ug/L	95
61) 1,2,3-Trichloropropane	11.740	75	13861	1.839	ug/L #	49
64) 1,3-Dichlorobenzene	11.740	146	38008	1.728	ug/L	96
65) 1,4-Dichlorobenzene	11.830	146	116416	5.354	ug/L	96
67) 1,2-Dichlorobenzene	12.206	146	4499	0.221	ug/L #	88
69) 1,3,5-Trichlorobenzene	13.849	180	2628	0.170	ug/L	94
70) 1,2,4-trichlorobenzene	13.849	180	2628	0.221	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100124\
Data File : VU061060.D
Acq On : 01 Oct 2024 21:53
Operator : MD/SY
Sample : P4227-07 400X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Oct 02 02:38:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 02 02:09:38 2024
Response via : Initial Calibration

