

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100124\
 Data File : VU061045.D
 Acq On : 01 Oct 2024 15:48
 Operator : MD/SY
 Sample : P4227-23
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB(20240924)

Quant Time: Oct 02 02:36:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 02 02:09:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	189784	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	176765	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	81892	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	63350	4.669	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.400%
7) Chloroethane-d5	1.911	69	47828	4.709	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.200%
11) 1,1-Dichloroethene-d2	2.563	65	27753	4.637	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.800%
20) 2-Butanone-d5	4.631	46	154823	47.635	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.280%
24) Chloroform-d	5.055	84	112924	4.722	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
26) 1,2-Dichloroethane-d4	5.695	65	62275	4.913	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.200%
32) Benzene-d6	5.721	84	229258	4.919	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.400%
36) 1,2-Dichloropropane-d6	6.682	67	75958	4.962	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.200%
41) Toluene-d8	7.891	98	210551	4.833	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.600%
43) trans-1,3-Dichloroprop...	8.177	79	31207	4.768	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.400%
46) 2-Hexanone-d5	8.627	63	129771	47.906	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.820%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	54165	4.825	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	67556	5.159	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.200%
Target Compounds					Qvalue	
3) Chloromethane	1.518	50	2454	0.184	ug/L	97
16) Methylene chloride	3.042	84	4329	0.344	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100124\
Data File : VU061045.D
Acq On : 01 Oct 2024 15:48
Operator : MD/SY
Sample : P4227-23
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB(20240924)

Quant Time: Oct 02 02:36:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 02 02:09:38 2024
Response via : Initial Calibration

