

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100124\
 Data File : VU061047.D
 Acq On : 01 Oct 2024 16:37
 Operator : MD/SY
 Sample : P4227-02 100X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PMW-2(20240925)

Quant Time: Oct 02 02:36:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 02 02:09:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	184773	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	173275	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	80244	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	58376	4.419	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.400%
7) Chloroethane-d5	1.911	69	45291	4.580	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.600%
11) 1,1-Dichloroethene-d2	2.563	65	25998	4.461	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.200%
20) 2-Butanone-d5	4.631	46	158043	49.945	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.880%
24) Chloroform-d	5.055	84	108999	4.682	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.600%
26) 1,2-Dichloroethane-d4	5.695	65	59508	4.822	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
32) Benzene-d6	5.718	84	219785	4.811	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.200%
36) 1,2-Dichloropropane-d6	6.682	67	72781	4.851	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.000%
41) Toluene-d8	7.891	98	202717	4.747	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.000%
43) trans-1,3-Dichloroprop...	8.174	79	30021	4.680	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.600%
46) 2-Hexanone-d5	8.628	63	129009	48.583	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.160%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	52511	4.772	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	63993	4.987	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.800%
Target Compounds						
5) Vinyl chloride	1.602	62	3666	0.281	ug/L #	79
12) 1,1-Dichloroethene	2.573	96	27115	2.541	ug/L #	47
16) Methylene chloride	3.043	84	2159	0.176	ug/L	93
18) trans-1,2-Dichloroethene	3.354	96	1170	0.100	ug/L	75
19) 1,1-Dichloroethane	3.859	63	375508	16.499	ug/L	98
22) cis-1,2-Dichloroethene	4.660	96	40030	3.073	ug/L	93
27) 1,2-Dichloroethane	5.801	62	3548	0.236	ug/L	100
34) Trichloroethene	6.534	95	70158	5.067	ug/L	96
65) 1,4-Dichlorobenzene	11.833	146	6875	0.302	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100124\
Data File : VU061047.D
Acq On : 01 Oct 2024 16:37
Operator : MD/SY
Sample : P4227-02 100X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PMW-2(20240925)

Quant Time: Oct 02 02:36:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 02 02:09:38 2024
Response via : Initial Calibration

