

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100218\
 Data File : VU027336.D
 Acq On : 02 Oct 2018 09:45
 Operator : MD/SY
 Sample : VU1002WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1002WBL01

Quant Time: Oct 03 02:11:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	100334	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	170620	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	161313	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	71954	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	102306	61.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	122.12%	
35) Dibromofluoromethane	4.88	113	71598	61.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	122.46%	
50) Toluene-d8	7.57	98	269684	61.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	123.24%	
62) 4-Bromofluorobenzene	10.31	95	86686	53.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.60%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100218\
Data File : VU027336.D
Acq On : 02 Oct 2018 09:45
Operator : MD/SY
Sample : VU1002WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1002WBL01

Quant Time: Oct 03 02:11:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

