

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100223\
 Data File : VU055607.D
 Acq On : 02 Oct 2023 16:14
 Operator : MD/SY
 Sample : 04679-02 200X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PMW-2(20230926)

Quant Time: Oct 03 04:57:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 27 23:33:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	152873	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	158036	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	76249	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	66447	4.301	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.000%
7) Chloroethane-d5	1.914	69	71602	5.259	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.200%
11) 1,1-Dichloroethene-d2	2.567	65	25165	3.957	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.200%
20) 2-Butanone-d5	4.637	46	142910	51.208	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.420%
24) Chloroform-d	5.062	84	131935	4.721	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
26) 1,2-Dichloroethane-d4	5.701	65	68133	5.068	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.400%
32) Benzene-d6	5.727	84	245233	4.436	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.800%
36) 1,2-Dichloropropane-d6	6.689	67	79125	4.791	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.800%
41) Toluene-d8	7.898	98	206547	4.100	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.000%
43) trans-1,3-Dichloroprop...	8.177	79	26604	4.590	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.800%
46) 2-Hexanone-d5	8.634	63	81891	46.450	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.900%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	62290	5.032	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.600%
66) 1,2-Dichlorobenzene-d4	12.193	152	72016	4.679	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.600%
Target Compounds						
5) Vinyl chloride	1.602	62	2173	0.154	ug/L #	33
12) 1,1-Dichloroethene	2.579	96	8524	0.797	ug/L #	1
16) Methylene chloride	3.042	84	10845	0.847	ug/L	90
19) 1,1-Dichloroethane	3.869	63	163477	8.205	ug/L	99
22) cis-1,2-Dichloroethene	4.669	96	17089	1.368	ug/L	93
34) Trichloroethene	6.541	95	25526	1.984	ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100223\
Data File : VU055607.D
Acq On : 02 Oct 2023 16:14
Operator : MD/SY
Sample : 04679-02 200X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PMW-2(20230926)

Quant Time: Oct 03 04:57:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 27 23:33:30 2023
Response via : Initial Calibration









#5
 Vinyl chloride
 Concen: 0.154 ug/L
 RT: 1.602 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: VU055607.D
 Acq: 02 Oct 2023 16:14




Instrument :
 MSVOA_U
 ClientSampleId :
 PMW-2(20230926)


Tgt Ion: 62 Resp: 2173
 Ion Ratio Lower Upper
 62 100
 64 69.3 22.3 41.5#



#12
 1,1-Dichloroethene
 Concen: 0.797 ug/L
 RT: 2.579 min Scan# 401
 Delta R.T. 0.003 min
 Lab File: VU055607.D
 Acq: 02 Oct 2023 16:14

Tgt Ion: 96 Resp: 8524
 Ion Ratio Lower Upper
 96 100
 61 211.6 114.5 212.7
 63 502.9 98.3 182.5#









#16
Methylene chloride
Concen: 0.847 ug/L
RT: 3.042 min Scan# 546
Delta R.T. -0.000 min
Lab File: VU055607.D
Acq: 02 Oct 2023 16:14




Instrument :
MSVOA_U
ClientSampleId :
PMW-2(20230926)


Tgt Ion: 84 Resp: 10845
Ion Ratio Lower Upper
84 100
86 66.3 44.9 83.3
49 126.9 78.8 146.3



#19
1,1-Dichloroethane
Concen: 8.205 ug/L
RT: 3.869 min Scan# 802
Delta R.T. 0.003 min
Lab File: VU055607.D
Acq: 02 Oct 2023 16:14

Tgt Ion: 63 Resp: 163477
Ion Ratio Lower Upper
63 100
65 32.6 22.0 41.0
83 14.3 10.1 18.9





#22

cis-1,2-Dichloroethene

Concen: 1.368 ug/L

RT: 4.669 min Scan# 1050

Delta R.T. 0.006 min

Lab File: VU055607.D

Acq: 02 Oct 2023 16:14

Instrument :

MSVOA_U

ClientSampleId :

PMW-2(20230926)





Tgt Ion: 96 Resp: 17089

Ion Ratio Lower Upper

96 100

61 115.5 87.8 163.2

98 65.5 44.4 82.4



#34

Trichloroethene

Concen: 1.984 ug/L

RT: 6.541 min Scan# 1633

Delta R.T. -0.000 min

Lab File: VU055607.D

Acq: 02 Oct 2023 16:14

Tgt Ion: 95 Resp: 25526

Ion Ratio Lower Upper

95 100

97 68.5 45.4 84.2

132 88.0 65.2 121.2

130 99.7 69.1 128.3

