

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100318\  
Data File : VU027358.D  
Acq On : 03 Oct 2018 12:12  
Operator : MD/SY  
Sample : VU1003WBL02  
Misc : 25.0mL/MSVOA U/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VU1003WBL02

Quant Time: Oct 03 14:42:21 2018  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U100318DW.M  
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
QLast Update : Wed Oct 03 09:31:21 2018  
Response via : Initial Calibration

| Internal Standards | R.T.  | QIon  | Response | Conc  | Units | Dev(Min) |
|--------------------|-------|-------|----------|-------|-------|----------|
| -----              | ----- | ----- | -----    | ----- | ----- | -----    |
| 1) Fluorobenzene   | 5.74  | 96    | 22023    | 1.00  | ug/l  | 0.00     |

## System Monitoring Compounds

|                            |       |     |          |      |        |      |
|----------------------------|-------|-----|----------|------|--------|------|
| 57) 4-Bromofluorobenzene   | 10.31 | 95  | 5865     | 0.70 | ug/l   | 0.00 |
| Spiked Amount 1.000        |       |     | Recovery | =    | 70.00% |      |
| 68) 1,2-Dichlorobenzene-d4 | 11.86 | 152 | 5038     | 0.71 | ug/l   | 0.00 |
| Spiked Amount 1.000        |       |     | Recovery | =    | 71.00% |      |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            | -----  |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100318\  
Data File : VU027358.D  
Acq On : 03 Oct 2018 12:12  
Operator : MD/SY  
Sample : VU1003WBL02  
Misc : 25.0mL/MSVOA U/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VU1003WBL02

Quant Time: Oct 03 14:42:21 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U100318DW.M  
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
QLast Update : Wed Oct 03 09:31:21 2018  
Response via : Initial Calibration

