

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U100320\
 Data File : VU040489.D
 Acq On : 03 Oct 2020 19:00
 Operator : SY/MD
 Sample : L4266-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Oct 05 02:53:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 05 00:56:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	121687	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	131640	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	59152	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	29556	4.91	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.20%
7) Chloroethane-d5	1.93	69	25712	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.40%
11) 1,1-Dichloroethene-d2	2.58	63	48468	3.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.00%
20) 2-Butanone-d5	4.66	46	132817	44.38	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.76%
24) Chloroform-d	5.08	84	64606	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.20%
26) 1,2-Dichloroethane-d4	5.72	65	41664	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
32) Benzene-d6	5.75	84	121261	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.60%
36) 1,2-Dichloropropane-d6	6.70	67	42957	4.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.80%
41) Toluene-d8	7.91	98	103942	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
43) trans-1,3-Dichloropropene-	8.19	79	17504	4.43	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.60%
46) 2-Hexanone-d5	8.64	63	109189	46.99	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.98%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	41230	4.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	43890	4.89	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.80%

Target Compounds

					Ovalue
3) Chloromethane	1.53	50	12436	1.039	ug/L 99
13) Acetone	2.67	43	5869	2.354	ug/L 81
34) Trichloroethene	6.56	95	30421	2.596	ug/L 94
42) Toluene	7.98	91	3335	0.074	ug/L 93
47) Tetrachloroethene	8.58	164	31206461m	3751.068	ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100320\
Data File : VU040489.D
Acq On : 03 Oct 2020 19:00
Operator : SY/MD
Sample : L4266-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Oct 05 02:53:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 05 00:56:13 2020
Response via : Initial Calibration

