

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U100320\
 Data File : VU040482.D
 Acq On : 03 Oct 2020 16:18
 Operator : SY/MD
 Sample : L4185-05
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK02

Quant Time: Oct 05 01:47:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 05 00:56:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	119445	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	121520	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	52856	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	30995	5.24	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.80%
7) Chloroethane-d5	1.92	69	26009	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.40%
11) 1,1-Dichloroethene-d2	2.58	63	49865	3.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.20%
20) 2-Butanone-d5	4.66	46	132462	45.10	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.20%
24) Chloroform-d	5.08	84	66597	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
26) 1,2-Dichloroethane-d4	5.72	65	41510	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
32) Benzene-d6	5.75	84	124790	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.20%
36) 1,2-Dichloropropane-d6	6.70	67	41963	4.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.00%
41) Toluene-d8	7.91	98	102085	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
43) trans-1,3-Dichloropropene-	8.19	79	16195	4.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.80%
46) 2-Hexanone-d5	8.64	63	86407	40.28	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.56%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	37767	4.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	40614	5.07	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.40%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	1945	0.170 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100320\
Data File : VU040482.D
Acq On : 03 Oct 2020 16:18
Operator : SY/MD
Sample : L4185-05
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK02

Quant Time: Oct 05 01:47:10 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 05 00:56:13 2020
Response via : Initial Calibration

