

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U100320\
 Data File : VU040493.D
 Acq On : 03 Oct 2020 20:33
 Operator : SY/MD
 Sample : L4266-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFSB9

Quant Time: Oct 05 03:07:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 05 00:56:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	117979	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	119335	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	51237	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	30815	5.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.60%
7) Chloroethane-d5	1.93	69	26707	5.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.40%
11) 1,1-Dichloroethene-d2	2.58	63	50982	3.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.60%
20) 2-Butanone-d5	4.66	46	140924	48.57	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.14%
24) Chloroform-d	5.08	84	68979	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
26) 1,2-Dichloroethane-d4	5.72	65	43407	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
32) Benzene-d6	5.74	84	129786	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.20%
36) 1,2-Dichloropropane-d6	6.70	67	44655	5.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.80%
41) Toluene-d8	7.91	98	108241	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.80%
43) trans-1,3-Dichloropropene-	8.19	79	17988	5.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.40%
46) 2-Hexanone-d5	8.64	63	94418	44.82	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.64%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	39434	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	42608	5.49	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.80%

Target Compounds

					Ovalue
3) Chloromethane	1.53	50	2944	0.254 ug/L	89
13) Acetone	2.67	43	10676	4.416 ug/L	71
42) Toluene	7.98	91	1540	0.038 ug/L	93
47) Tetrachloroethene	8.56	164	5732	0.760 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100320\
Data File : VU040493.D
Acq On : 03 Oct 2020 20:33
Operator : SY/MD
Sample : L4266-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFSB9

Quant Time: Oct 05 03:07:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR092920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 05 00:56:13 2020
Response via : Initial Calibration

