

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100323\
 Data File : VU055644.D
 Acq On : 03 Oct 2023 16:32
 Operator : MD/SY
 Sample : 04692-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCJK0

Quant Time: Oct 04 04:44:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 04 04:40:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	177007	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	183287	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	98335	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	86593	4.840	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.800%
7) Chloroethane-d5	1.911	69	69028	4.378	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.600%
11) 1,1-Dichloroethene-d2	2.567	65	31882	4.330	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.600%
20) 2-Butanone-d5	4.628	46	127728	39.528	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.060%
24) Chloroform-d	5.062	84	139757	4.319	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.400%
26) 1,2-Dichloroethane-d4	5.698	65	65880	4.232	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
32) Benzene-d6	5.724	84	280098	4.369	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.400%
36) 1,2-Dichloropropane-d6	6.689	67	84029	4.387	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.800%
41) Toluene-d8	7.898	98	242091	4.143	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.800%
43) trans-1,3-Dichloroprop...	8.181	79	28746	4.276	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.600%
46) 2-Hexanone-d5	8.634	63	72163	35.293	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.580%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	57892	4.033	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	79412	4.001	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	80.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100323\
Data File : VU055644.D
Acq On : 03 Oct 2023 16:32
Operator : MD/SY
Sample : 04692-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCJK0

Quant Time: Oct 04 04:44:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 04 04:40:48 2023
Response via : Initial Calibration

