

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100418\
 Data File : VU027370.D
 Acq On : 03 Oct 2018 21:06
 Operator : MD/SY
 Sample : J5155-04DL 250X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 QNWP8-MW9-GWDL

Quant Time: Oct 04 00:25:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	95867	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	166397	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	158151	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	83258	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	100960	63.07	ug/l	0.00
Spiked Amount			Recovery	=	126.14%	
35) Dibromofluoromethane	4.89	113	70371	61.71	ug/l	0.00
Spiked Amount			Recovery	=	123.42%	
50) Toluene-d8	7.57	98	255801	59.93	ug/l	0.00
Spiked Amount			Recovery	=	119.86%	
62) 4-Bromofluorobenzene	10.31	95	89123	56.19	ug/l	0.00
Spiked Amount			Recovery	=	112.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
40) Benzene	5.39	78	20534	3.902	ug/l	99
52) Toluene	7.64	92	252969	81.211	ug/l	99
67) Ethyl Benzene	9.25	91	22574	3.827	ug/l	100
68) m/p-Xylenes	9.38	106	23357	10.826	ug/l	97
69) o-Xylene	9.78	106	10241	4.995	ug/l	95
80) 1,3,5-Trimethylbenzene	10.77	105	3046	0.619	ug/l	98
84) 1,2,4-Trimethylbenzene	11.15	105	7820	1.555	ug/l	96
95) Naphthalene	13.74	128	347757	55.073	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100418\
Data File : VU027370.D
Acq On : 03 Oct 2018 21:06
Operator : MD/SY
Sample : J5155-04DL 250X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
QNWP8-MW9-GWDL

Quant Time: Oct 04 00:25:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

