

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100418\
 Data File : VU027371.D
 Acq On : 03 Oct 2018 21:29
 Operator : MD/SY
 Sample : J5155-05DL 500X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 QNWP8-MW16S-GWDL

Quant Time: Oct 05 17:56:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	93193	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	164617	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	154362	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	74572	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	100087	64.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.64%	
35) Dibromofluoromethane	4.88	113	68382	60.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.22%	
50) Toluene-d8	7.57	98	254412	60.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.50%	
62) 4-Bromofluorobenzene	10.31	95	82815	52.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.56%	
Target Compounds						
40) Benzene	5.39	78	2367	0.455	ug/l	96
67) Ethyl Benzene	9.25	91	12231	2.125	ug/l	98
68) m/p-Xylenes	9.38	106	2691	1.278	ug/l	96
69) o-Xylene	9.78	106	1353	0.676	ug/l	86
80) 1,3,5-Trimethylbenzene	10.77	105	1876	0.425	ug/l	97
84) 1,2,4-Trimethylbenzene	11.15	105	4577	1.016	ug/l	95
95) Naphthalene	13.75	128	229447	41.006	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100418\
Data File : VU027371.D
Acq On : 03 Oct 2018 21:29
Operator : MD/SY
Sample : J5155-05DL 500X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
QNWP8-MW16S-GWDL

Quant Time: Oct 05 17:56:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

