

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100418\
 Data File : VU027375.D
 Acq On : 03 Oct 2018 23:02
 Operator : MD/SY
 Sample : J5163-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-3

Quant Time: Oct 04 00:58:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	90073	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	161976	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	149604	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	70137	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	96910	64.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.86%	
35) Dibromofluoromethane	4.89	113	68639	61.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.66%	
50) Toluene-d8	7.57	98	247066	59.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.94%	
62) 4-Bromofluorobenzene	10.31	95	78869	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
Target Compounds						
16) Acetone	2.32	43	3944	4.439	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100418\
Data File : VU027375.D
Acq On : 03 Oct 2018 23:02
Operator : MD/SY
Sample : J5163-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-3

Quant Time: Oct 04 00:58:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

