

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100418\
 Data File : VU027380.D
 Acq On : 04 Oct 2018 00:58
 Operator : MD/SY
 Sample : J5163-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-2R

Quant Time: Oct 04 08:14:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	92120	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	170977	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	156418	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	79147	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	97866	63.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.24%	
35) Dibromofluoromethane	4.89	113	69374	59.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.42%	
50) Toluene-d8	7.57	98	265152	60.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.92%	
62) 4-Bromofluorobenzene	10.31	95	85498	52.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.92%	
Target Compounds						
11) Tert butyl alcohol	2.83	59	848397	2619.086	ug/l	98
16) Acetone	2.32	43	5505	6.058	ug/l	97
19) Methyl tert-butyl Ether	3.00	73	1244501	319.193	ug/l	99
85) sec-Butylbenzene	11.32	105	9716	1.704	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100418\
Data File : VU027380.D
Acq On : 04 Oct 2018 00:58
Operator : MD/SY
Sample : J5163-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-2R

Quant Time: Oct 04 08:14:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

