

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100418\
 Data File : VU027384.D
 Acq On : 04 Oct 2018 02:31
 Operator : MD/SY
 Sample : J5218-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 280-356

Quant Time: Oct 04 08:24:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	91682	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	167685	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	163312	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	83311	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	101520	66.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	132.62%	
35) Dibromofluoromethane	4.88	113	69102	60.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.26%	
50) Toluene-d8	7.57	98	258051	60.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.00%	
62) 4-Bromofluorobenzene	10.31	95	87152	54.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.04%	
Target Compounds						
					Qvalue	
16) Acetone	2.32	43	7293	8.064	ug/l	100
18) Methyl Acetate	2.62	43	2221	1.076	ug/l	94
20) Methylene Chloride	2.70	84	3924	3.115	ug/l	96
30) Chloroform	4.68	83	2338	1.095	ug/l	95
43) Isopropyl Acetate	5.55	43	86352	23.629	ug/l	99
95) Naphthalene	13.75	128	100465	17.081	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100418\
Data File : VU027384.D
Acq On : 04 Oct 2018 02:31
Operator : MD/SY
Sample : J5218-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
280-356

Quant Time: Oct 04 08:24:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

