

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U100418\
 Data File : VU027405.D
 Acq On : 04 Oct 2018 13:55
 Operator : MD/SY
 Sample : J5084-07DL 100X
 Misc : 4.90uL/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 B-10(13-13.5)DL

Quant Time: Oct 05 04:38:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	102797	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	181181	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	174636	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	96835	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	101030	58.86	ug/l	0.00
Spiked Amount			Recovery	=	117.72%	
35) Dibromofluoromethane	4.88	113	71280	57.41	ug/l	0.00
Spiked Amount			Recovery	=	114.82%	
50) Toluene-d8	7.57	98	283943	61.10	ug/l	0.00
Spiked Amount			Recovery	=	122.20%	
62) 4-Bromofluorobenzene	10.31	95	103599	59.98	ug/l	0.00
Spiked Amount			Recovery	=	119.96%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
39) Methylcyclohexane	6.42	83	10126	4.403	ug/l	98
52) Toluene	7.64	92	37517	11.061	ug/l	98
67) Ethyl Benzene	9.25	91	83854	12.875	ug/l	99
68) m/p-Xylenes	9.38	106	83114	34.889	ug/l	100
69) o-Xylene	9.78	106	22719	10.036	ug/l	100
73) Isopropylbenzene	10.17	105	8258	1.174	ug/l	99
78) n-propylbenzene	10.59	91	41302	4.887	ug/l	99
80) 1,3,5-Trimethylbenzene	10.77	105	40491	7.072	ug/l	99
84) 1,2,4-Trimethylbenzene	11.15	105	139408	23.835	ug/l	100
85) sec-Butylbenzene	11.33	105	4380	0.628	ug/l	93
86) p-Isopropyltoluene	11.43	119	5176	0.867	ug/l	92
89) n-Butylbenzene	11.89	91	16214	4.408	ug/l #	47
95) Naphthalene	13.75	128	22984	4.695	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100418\
Data File : VU027405.D
Acq On : 04 Oct 2018 13:55
Operator : MD/SY
Sample : J5084-07DL 100X
Misc : 4.90µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
B-10(13-13.5)DL

Quant Time: Oct 05 04:38:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

