

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100418\
 Data File : VU027409.D
 Acq On : 04 Oct 2018 15:28
 Operator : MD/SY
 Sample : J5165-01
 Misc : 6.10g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EP1-MW-G-092718(7-10)

Quant Time: Oct 05 04:50:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	103481	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	175720	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	184719	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	115207	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	101880	58.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.92%	
35) Dibromofluoromethane	4.88	113	70860	58.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.68%	
50) Toluene-d8	7.57	98	275961	61.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.46%	
62) 4-Bromofluorobenzene	10.31	95	127689	76.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	152.46%	
Target Compounds						
73) Isopropylbenzene	10.17	105	18295	2.186	ug/l	Qvalue 100
78) n-propylbenzene	10.59	91	80083	7.965	ug/l	98
83) tert-Butylbenzene	11.10	119	8773	1.332	ug/l	82
84) 1,2,4-Trimethylbenzene	11.15	105	135925	19.533	ug/l	99
85) sec-Butylbenzene	11.33	105	74203	8.943	ug/l	99
89) n-Butylbenzene	11.89	91	59407	9.522	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100418\
Data File : VU027409.D
Acq On : 04 Oct 2018 15:28
Operator : MD/SY
Sample : J5165-01
Misc : 6.10µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EP1-MW-G-092718(7-10)

Quant Time: Oct 05 04:50:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

