

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100418\
 Data File : VU027422.D
 Acq On : 04 Oct 2018 20:47
 Operator : MD/SY
 Sample : J5255-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-2

Quant Time: Oct 08 11:01:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	109538	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	189237	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	181390	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	98039	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	105402	57.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.26%	
35) Dibromofluoromethane	4.89	113	73660	56.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.60%	
50) Toluene-d8	7.57	98	285498	58.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.64%	
62) 4-Bromofluorobenzene	10.31	95	104383	57.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.74%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100418\
Data File : VU027422.D
Acq On : 04 Oct 2018 20:47
Operator : MD/SY
Sample : J5255-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-2

Quant Time: Oct 08 11:01:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

