

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU100418\
 Data File : VU027392.D
 Acq On : 04 Oct 2018 08:45
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 05 03:43:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.721	0.731	-1.4	86	0.00
3 P	Chloromethane	0.976	0.844	13.5	79	0.00
4 C	Vinyl Chloride	0.917	0.874	4.7#	86	0.00
5 T	Bromomethane	0.397	0.432	-8.8	106	0.00
6 T	Chloroethane	0.551	0.492	10.7	86	0.00
7 T	Trichlorofluoromethane	0.931	0.931	0.0	91	0.00
8 T	Diethyl Ether	0.429	0.400	6.8	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.541	0.517	4.4	87	0.00
10 T	Methyl Iodide	0.394	0.459	-16.5	102	0.00
11 T	Tert butyl alcohol	0.176	0.159	9.7	87	0.00
12 CM	1,1-Dichloroethene	0.528	0.476	9.8#	82	0.00
13 T	Acrolein	0.148	0.149	-0.7	99	0.00
14 T	Allyl chloride	1.254	1.186	5.4	86	0.00
15 T	Acrylonitrile	0.458	0.427	6.8	85	0.00
16 T	Acetone	0.493	0.417	15.4	79	0.00
17 T	Carbon Disulfide	1.905	1.679	11.9	83	0.00
18 T	Methyl Acetate	1.125	1.114	1.0	94	0.00
19 T	Methyl tert-butyl Ether	2.116	2.021	4.5	88	0.00
20 T	Methylene Chloride	0.687	0.643	6.4	88	0.00
21 T	trans-1,2-Dichloroethene	0.593	0.560	5.6	88	0.00
22 T	Diisopropyl ether	2.549	2.510	1.5	89	0.00
23 T	Vinyl Acetate	2.226	2.143	3.7	85	0.00
24 P	1,1-Dichloroethane	1.310	1.275	2.7	88	0.00
25 T	2-Butanone	0.673	0.616	8.5	82	0.00
26 T	2,2-Dichloropropane	1.077	1.014	5.8	88	0.00
27 T	cis-1,2-Dichloroethene	0.678	0.637	6.0	87	0.00
28 T	Bromochloromethane	0.632	0.648	-2.5	91	0.00
29 T	Tetrahydrofuran	0.424	0.393	7.3	83	0.00
30 C	Chloroform	1.164	1.151	1.1#	90	0.00
31 T	Cyclohexane	1.532	1.195	22.0#	84	0.00
32 T	1,1,1-Trichloroethane	0.981	0.948	3.4	88	0.00
33 S	1,2-Dichloroethane-d4	0.835	0.761	8.9	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.343	0.326	5.0	93	0.00
36 T	1,1-Dichloropropene	0.543	0.528	2.8	87	0.00
37 T	Ethyl Acetate	0.702	0.674	4.0	86	0.00
38 T	Carbon Tetrachloride	0.490	0.486	0.8	89	0.00
39 T	Methylcyclohexane	0.635	0.619	2.5	86	0.00
40 TM	Benzene	1.581	1.581	0.0	89	0.00
41 T	Methacrylonitrile	0.397	0.386	2.8	85	0.00
42 TM	1,2-Dichloroethane	0.587	0.592	-0.9	89	0.00
43 T	Isopropyl Acetate	1.090	1.077	1.2	85	0.00
44 TM	Trichloroethene	0.353	0.342	3.1	87	0.00
45 C	1,2-Dichloropropane	0.456	0.454	0.4#	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU100418\
 Data File : VU027392.D
 Acq On : 04 Oct 2018 08:45
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 05 03:43:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.258	0.264	-2.3	89	0.00
47 T	Bromodichloromethane	0.535	0.534	0.2	87	0.00
48 T	Methyl methacrylate	0.542	0.543	-0.2	85	0.00
49 T	1,4-Dioxane	0.009	0.010	-11.1	91	0.00
50 S	Toluene-d8	1.282	1.256	2.0	91	0.00
51 T	4-Methyl-2-Pentanone	0.708	0.703	0.7	85	0.00
52 CM	Toluene	0.936	0.925	1.2#	87	0.00
53 T	t-1,3-Dichloropropene	0.627	0.630	-0.5	87	0.00
54 T	cis-1,3-Dichloropropene	0.674	0.682	-1.2	87	0.00
55 T	1,1,2-Trichloroethane	0.371	0.365	1.6	86	0.00
56 T	Ethyl methacrylate	0.616	0.620	-0.6	83	0.00
57 T	1,3-Dichloropropane	0.720	0.711	1.3	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.344	0.350	-1.7	85	0.00
59 T	2-Hexanone	0.549	0.544	0.9	85	0.00
60 T	Dibromochloromethane	0.372	0.380	-2.2	90	0.00
61 T	1,2-Dibromoethane	0.395	0.384	2.8	85	0.00
62 S	4-Bromofluorobenzene	0.477	0.465	2.5	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.311	0.307	1.3	86	0.00
65 PM	Chlorobenzene	1.052	1.044	0.8	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.368	-1.4	88	0.00
67 C	Ethyl Benzene	1.865	1.890	-1.3#	86	0.00
68 T	m/p-Xylenes	0.682	0.710	-4.1	88	0.00
69 T	o-Xylene	0.648	0.674	-4.0	86	0.00
70 T	Styrene	1.064	1.156	-8.6	89	0.00
71 P	Bromoform	0.292	0.295	-1.0	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.631	3.479	4.2	87	0.00
74 T	N-amyl acetate	2.079	1.945	6.4	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.426	1.297	9.0	88	0.00
76 T	1,2,3-Trichloropropane	1.296	1.114	14.0	81	0.00
77 T	Bromobenzene	0.844	0.797	5.6	89	0.00
78 T	n-propylbenzene	4.364	4.375	-0.3	90	0.00
79 T	2-Chlorotoluene	2.635	2.504	5.0	89	0.00
80 T	1,3,5-Trimethylbenzene	2.957	2.988	-1.0	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.495	0.445	10.1	90	0.00
82 T	4-Chlorotoluene	2.990	2.962	0.9	90	0.00
83 T	tert-Butylbenzene	2.858	2.796	2.2	88	0.00
84 T	1,2,4-Trimethylbenzene	3.020	3.065	-1.5	91	0.00
85 T	sec-Butylbenzene	3.601	3.654	-1.5	91	0.00
86 T	p-Isopropyltoluene	3.083	3.167	-2.7	91	0.00
87 T	1,3-Dichlorobenzene	1.598	1.515	5.2	90	0.00
88 T	1,4-Dichlorobenzene	1.633	1.544	5.5	91	0.00
89 T	n-Butylbenzene	2.876	3.156	-9.7	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU100418\
Data File : VU027392.D
Acq On : 04 Oct 2018 08:45
Operator : MD/SY
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Oct 05 03:43:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.581	0.556	4.3	89	0.00
91 T	1,2-Dichlorobenzene	1.556	1.482	4.8	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.316	0.292	7.6	85	0.00
93 T	1,2,4-Trichlorobenzene	1.007	1.091	-8.3	97	0.00
94 T	Hexachlorobutadiene	0.509	0.539	-5.9	99	0.00
95 T	Naphthalene	3.375	3.510	-4.0	92	0.00
96 T	1,2,3-Trichlorobenzene	1.049	1.110	-5.8	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6