

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U100419\
 Data File : VU034941.D
 Acq On : 04 Oct 2019 17:38
 Operator : JC/SP
 Sample : K5201-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AC0

Quant Time: Oct 05 04:09:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR093019WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 05 00:29:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	196247	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	194161	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.46	152	83371	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	73091	5.26	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.20%
7) Chloroethane-d5	1.67	69	61233	5.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.00%
11) 1,1-Dichloroethene-d2	2.26	63	99119	4.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.20%
20) 2-Butanone-d5	4.17	46	173057	43.75	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.50%
24) Chloroform-d	4.62	84	123592	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.20%
26) 1,2-Dichloroethane-d4	5.29	65	67614	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.60%
32) Benzene-d6	5.32	84	254498	5.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.40%
36) 1,2-Dichloropropane-d6	6.31	67	85621	5.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.20%
41) Toluene-d8	7.54	98	230587	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.20%
43) trans-1,3-Dichloropropene-	7.84	79	26932	4.58	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.60%
46) 2-Hexanone-d5	8.30	63	135587	46.71	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.42%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	66110	5.19	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.80%
64) 1,2-Dichlorobenzene-d4	11.84	152	81835	5.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.40%

Target Compounds					Ovalue
13) Acetone	2.31	43	4809	1.661 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100419\
Data File : VU034941.D
Acq On : 04 Oct 2019 17:38
Operator : JC/SP
Sample : K5201-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AC0

Quant Time: Oct 05 04:09:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR093019WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 05 00:29:19 2019
Response via : Initial Calibration

