

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100523\
 Data File : VU055677.D
 Acq On : 05 Oct 2023 17:06
 Operator : MD/SY
 Sample : 04744-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH583

Quant Time: Oct 06 07:11:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100523.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 06 06:49:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	9061	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	14959	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	4932	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	123	0.022	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	4.000%#
4) 1,2-Dichloroethane-d4	5.685	65	3707	0.545	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	108.000%
7) 1,2-Dichloropropane-d6	6.687	67	4429	0.503	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	100.000%
8) Toluene-d8	7.887	98	2146	0.108	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	22.000%#
10) 1,1,2,2-Tetrachloroeth...	10.745	84	3667	0.523	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	104.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100523\
Data File : VU055677.D
Acq On : 05 Oct 2023 17:06
Operator : MD/SY
Sample : 04744-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH583

Quant Time: Oct 06 07:11:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100523.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 06 06:49:44 2023
Response via : Initial Calibration

