

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100523\
 Data File : VU055674.D
 Acq On : 05 Oct 2023 15:53
 Operator : MD/SY
 Sample : 04744-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH579

Quant Time: Oct 06 07:10:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100523.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 06 06:49:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	10152	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	15218	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	5494	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	975	0.159	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	32.000%#
4) 1,2-Dichloroethane-d4	5.685	65	4830	0.634	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	126.000%
7) 1,2-Dichloropropane-d6	6.687	67	6000	0.670	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	134.000%
8) Toluene-d8	7.888	98	9312	0.461	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.000%
10) 1,1,2,2-Tetrachloroeth...	10.746	84	4848	0.680	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	136.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100523\
Data File : VU055674.D
Acq On : 05 Oct 2023 15:53
Operator : MD/SY
Sample : 04744-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH579

Quant Time: Oct 06 07:10:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100523.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 06 06:49:44 2023
Response via : Initial Calibration

